
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005434
11:56 25/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:57
STERILIZING 2:59
136.0 C 220kPa 2:59
136.4 C 224kPa 3:29
136.3 C 224kPa 3:59
136.3 C 223kPa 4:29
136.4 C 223kPa 4:59
136.4 C 222kPa 5:29
136.4 C 221kPa 5:59
136.4 C 224kPa 6:29
136.4 C 224kPa 6:29
VENTING 6:30
AIR DRYING 7:23
CYCLE COMPLETE 10:01

Digital Signature #
D2279399C2654DCE
