
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001285
8:27 26/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:46
STERILIZING 4:49
136.0 C 220kPa 4:49
136.4 C 223kPa 5:19
136.4 C 224kPa 5:49
136.3 C 221kPa 6:19
136.3 C 223kPa 6:49
136.4 C 224kPa 7:19
136.1 C 222kPa 7:49
136.4 C 222kPa 8:19
136.4 C 222kPa 8:19
VENTING 8:19
AIR DRYING 9:13
CYCLE COMPLETE 9:22

Digital Signature #
C68573DE5B02246B
