
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002168
7:18 04/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:32
STERILIZING 3:35
136.4 C 220kPa 3:35
136.4 C 221kPa 4:05
136.4 C 219kPa 4:35
135.9 C 220kPa 5:05
136.2 C 222kPa 5:35
136.4 C 223kPa 6:05
136.4 C 223kPa 6:35
136.2 C 223kPa 7:05
136.2 C 223kPa 7:05
VENTING 7:05
AIR DRYING 7:56
CYCLE COMPLETE 8:02

Digital Signature #
C2965F3B3FFC55DC
