
Midmark M11 - v1.0.5
Total Cycles: 2740

Name: RITTER2
Sterilizer ID: V2313228
08/08/2023 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 190.6 F 3.2
4:00 205.0 F 9.2
6:00 221.0 F 6.5
8:00 236.3 F 11.6
10:00 237.0 F 9.9
12:00 236.0 F 8.8
14:00 252.0 F 16.6
16:00 264.7 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 271.9 F 28.7
2:00 272.4 F 29.0
3:00 272.8 F 29.5
4:00 272.0 F 29.2

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 201.9 F 0.0
10:00 218.2 F 0.0
15:00 221.3 F 0.0
20:00 221.0 F 0.0
25:00 218.3 F 0.0
30:00 214.2 F 0.0

Drying Complete

FILLING: 2:37
HEATING: 17:41
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:57:15
CYCLE COMPLETE

08/08/2023 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____