
Midmark M11 - v1.0.5
Total Cycles: 1602

Name: RITTER2
Sterilizer ID: V2313228
10/14/2022 04:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 186.0 F 3.7
4:00 199.2 F 8.8
6:00 216.5 F 7.9
8:00 231.0 F 9.9
10:00 234.6 F 8.6
12:00 243.2 F 11.9
14:00 246.5 F 13.7
16:00 259.7 F 20.9
18:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.1
2:00 272.1 F 29.2
3:00 272.1 F 28.9
4:00 272.0 F 29.2

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.1 F 0.0
5:00 223.8 F 0.0
10:00 235.6 F 0.0
15:00 229.9 F 0.0
20:00 224.0 F 0.0
25:00 222.6 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

10/14/2022 05:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____