

-----  
Midmark M11 - v1.0.5  
Total Cycles: 234

Name: RITTER2  
Sterilizer ID: V2313228  
07/21/2021 02:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 167.2 F 3.2  
4:00 189.1 F 7.5  
6:00 210.1 F 12.4  
8:00 223.3 F 6.8  
10:00 237.1 F 12.2  
12:00 239.1 F 10.9  
14:00 239.3 F 10.5  
16:00 253.5 F 17.5  
18:00 265.2 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.6 F 29.3  
2:00 271.5 F 29.1  
3:00 272.6 F 29.4  
4:00 271.6 F 28.8  
  
Min 271.4 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.8 F 0.0  
5:00 243.1 F 0.0  
10:00 233.0 F 0.0  
Dry Cycle Aborted

FILLING: 1:46  
HEATING: 19:45  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 12:57  
TOTAL CYCLE: 00:41:47

07/21/2021 03:14 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01