

-----  
Midmark M11 - v1.0.5  
Total Cycles: 250

Name: RITTER2  
Sterilizer ID: V2313228  
07/27/2021 07:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.8 F 0.0  
2:00 120.7 F 0.9  
4:00 139.0 F 3.4  
6:00 164.4 F 6.8  
8:00 186.7 F 11.0  
10:00 209.6 F 4.1  
12:00 221.2 F 8.1  
14:00 232.7 F 12.6  
16:00 234.9 F 9.6  
18:00 237.8 F 9.8  
20:00 246.3 F 14.0  
22:00 258.0 F 20.2  
24:00 268.2 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 271.7 F 29.2  
2:00 271.5 F 29.1  
3:00 271.9 F 29.3  
4:00 272.3 F 29.6

Min 271.4 F 29.1  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.6 F 0.0  
5:00 202.7 F 0.0  
10:00 204.0 F 0.0  
15:00 205.3 F 0.0  
20:00 197.0 F 0.0  
25:00 194.4 F 0.0  
30:00 206.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 25:09  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:03:23  
CYCLE COMPLETE

07/27/2021 09:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01