
Midmark M11 - v1.0.5
Total Cycles: 3650

Name: RITTER2
Sterilizer ID: V2313228
05/28/2024 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 69.8 F 0.0
4:00 79.4 F 0.2
6:00 99.3 F 1.6
8:00 133.9 F 4.2
10:00 165.1 F 7.4
12:00 206.8 F 11.0
14:00 218.8 F 4.1
16:00 219.5 F 6.6
18:00 229.7 F 10.4
20:00 232.7 F 8.8
22:00 239.2 F 11.1
24:00 234.7 F 8.7
26:00 248.2 F 14.7
28:00 258.9 F 20.7
30:00 268.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.8
2:00 272.4 F 29.8
3:00 272.7 F 29.8
4:00 272.7 F 29.8

Min 270.6 F 28.7
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 227.8 F 0.0
10:00 190.8 F 0.0
15:00 175.4 F 0.0
20:00 173.6 F 0.0
25:00 171.5 F 0.0
30:00 169.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:03
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

05/28/2024 09:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____