

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3442

Name: RITTER 1  
Sterilizer ID: V2178198  
06/26/2023 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.9 F 0.0  
2:00 175.7 F 0.2  
4:00 171.0 F 2.1  
6:00 193.5 F 5.8  
8:00 212.4 F 10.6  
10:00 221.3 F 5.5  
12:00 234.8 F 10.6  
14:00 232.8 F 7.5  
16:00 244.7 F 12.8  
18:00 244.1 F 12.4  
20:00 256.9 F 19.0  
22:00 267.9 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.0 F 28.9  
2:00 272.0 F 28.6  
3:00 271.5 F 28.6  
4:00 272.5 F 29.0

Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 225.6 F 0.0  
10:00 225.7 F 0.0  
15:00 225.9 F 0.0  
20:00 215.3 F 0.0  
25:00 214.2 F 0.0  
30:00 213.3 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 23:11  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:18  
CYCLE COMPLETE

06/26/2023 10:11 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_