
Midmark M11 - v1.0.5
Total Cycles: 4243

Name: RITTER2
Sterilizer ID: V2313228
01/06/2025 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.0 F 0.0
2:00 101.4 F 0.0
4:00 122.0 F 1.1
6:00 151.7 F 3.8
8:00 173.7 F 7.0
10:00 192.2 F 10.7
12:00 205.6 F 5.1
14:00 217.7 F 6.4
16:00 228.6 F 10.4
18:00 232.0 F 8.3
20:00 242.0 F 12.5
22:00 240.8 F 11.2
24:00 253.0 F 17.4
26:00 263.9 F 23.8
28:00 271.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.4 F 29.2
2:00 272.6 F 29.8
3:00 271.9 F 28.9
4:00 272.9 F 29.5

Min 271.8 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.0 F 0.0
5:00 216.6 F 0.0
10:00 197.0 F 0.0
15:00 193.9 F 0.0
20:00 188.8 F 0.0
25:00 192.7 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:05
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:06:36
CYCLE COMPLETE

01/06/2025 02:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____