

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2493

Name: RITTER2  
Sterilizer ID: V2313228  
06/09/2023 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
2:00 183.2 F 2.8  
4:00 185.3 F 7.3  
6:00 204.6 F 12.6  
8:00 227.7 F 8.1  
10:00 240.8 F 12.2  
12:00 240.0 F 11.6  
14:00 238.9 F 10.5  
16:00 252.7 F 17.3  
18:00 264.5 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.8 F 30.0  
2:00 271.9 F 29.8  
3:00 271.8 F 29.3  
4:00 272.8 F 29.7  
  
Min 270.8 F 28.8  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 230.0 F 0.0  
10:00 234.8 F 0.0  
15:00 228.6 F 0.0  
20:00 228.5 F 0.0  
25:00 229.3 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 1:42  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:48  
CYCLE COMPLETE

06/09/2023 11:00 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_