
Midmark M11 - v1.0.5
Total Cycles: 4330

Name: RITTER2
Sterilizer ID: V2313228
02/04/2025 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.6 F 0.1
4:00 83.9 F 1.2
6:00 114.8 F 3.7
8:00 147.4 F 7.2
10:00 173.6 F 11.3
12:00 210.9 F 2.5
14:00 214.7 F 6.7
16:00 227.5 F 10.8
18:00 232.2 F 7.2
20:00 239.1 F 11.7
22:00 237.6 F 9.4
24:00 250.9 F 15.9
26:00 262.7 F 22.8
28:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 271.8 F 28.7
2:00 272.0 F 28.7
3:00 272.1 F 28.9
4:00 272.2 F 29.0

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 230.7 F 0.0
Dry Cycle Aborted

FILLING: 0:42
HEATING: 28:13
STERILIZING: 4:00
VENTING: 2:58
DRYING: 0:13
TOTAL CYCLE: 00:36:06

02/04/2025 09:58 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____