
Midmark M11 - v1.0.0
Total Cycles: 4657

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2024 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 168.2 F 2.3
4:00 171.7 F 5.7
6:00 192.3 F 9.2
8:00 214.9 F 9.8
10:00 219.3 F 6.5
12:00 229.6 F 10.8
14:00 230.5 F 7.1
16:00 242.1 F 12.2
18:00 240.0 F 10.4
20:00 252.1 F 16.3
22:00 262.6 F 22.4
24:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 272.0 F 28.5
3:00 271.5 F 28.4
4:00 272.6 F 29.1

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 213.9 F 0.0
10:00 220.5 F 0.0
15:00 223.5 F 0.0
20:00 229.1 F 0.0
25:00 228.7 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:31
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

06/19/2024 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____