
Midmark M11 - v1.0.5
Total Cycles: 553

Name: RITTER2
Sterilizer ID: V2313228
11/15/2021 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.1 F 0.0
2:00 120.2 F 1.4
4:00 153.7 F 4.5
6:00 180.9 F 8.7
8:00 200.7 F 13.2
10:00 216.4 F 7.3
12:00 230.2 F 12.3
14:00 232.6 F 9.4
16:00 236.5 F 9.7
18:00 246.7 F 14.2
20:00 258.6 F 20.8
22:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.6
2:00 271.4 F 29.4
3:00 272.4 F 29.6
4:00 271.0 F 29.1

Min 271.0 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 230.2 F 0.0
10:00 220.3 F 0.0
15:00 219.8 F 0.0
20:00 206.5 F 0.0
25:00 203.3 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:40
CYCLE COMPLETE

11/15/2021 04:06 PM
ACCEPTED: X
REJECTED:
Operator: U02