
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000271
9:49 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 5:10
136.4 C 234kPa 5:10
136.2 C 232kPa 5:40
136.2 C 232kPa 6:10
136.2 C 232kPa 6:40
136.4 C 232kPa 7:10
136.4 C 233kPa 7:40
136.4 C 233kPa 8:10
136.3 C 232kPa 8:40
136.3 C 232kPa 8:40
VENTING 8:40
AIR DRYING 9:42
CYCLE COMPLETE 24:42

Digital Signature #
CA259B10C986F1D8
