
Midmark M11 - v1.0.5
Total Cycles: 1986

Name: RITTER2
Sterilizer ID: V2313228
01/31/2023 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.7 F 0.0
2:00 185.4 F 3.7
4:00 189.6 F 8.5
6:00 210.0 F 8.6
8:00 226.7 F 9.3
10:00 234.0 F 9.2
12:00 243.4 F 13.2
14:00 243.6 F 12.3
16:00 256.5 F 19.2
18:00 268.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.5
2:00 271.6 F 28.8
3:00 271.7 F 29.4
4:00 272.3 F 29.3

Min 271.1 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 218.9 F 0.0
10:00 224.6 F 0.0
15:00 232.0 F 0.0
20:00 226.3 F 0.0
25:00 225.8 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

01/31/2023 01:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____