
Midmark M11 - v1.0.0
Total Cycles: 1491

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 04:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.6 F 0.0
2:00 198.9 F 3.7
4:00 204.7 F 8.8
6:00 222.7 F 9.0
8:00 231.5 F 10.1
10:00 233.4 F 8.0
12:00 244.5 F 12.7
14:00 244.7 F 12.6
16:00 257.4 F 19.5
18:00 268.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.6
3:00 272.4 F 29.1
4:00 271.5 F 28.4

Min 271.3 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 222.1 F 0.0
10:00 233.6 F 0.0
15:00 232.5 F 0.0
20:00 225.1 F 0.0
25:00 219.3 F 0.0
30:00 219.4 F 0.0

Drying Complete

FILLING: 2:25
HEATING: 19:07
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

12/30/2021 07:58 AM
ACCEPTED: X
REJECTED:
Operator: U02