
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000153
11:23 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:57
STERILIZING 3:52
136.4 C 220kPa 3:52
136.2 C 224kPa 4:23
136.4 C 223kPa 4:53
136.1 C 221kPa 5:23
136.4 C 224kPa 5:53
136.3 C 223kPa 6:23
136.4 C 223kPa 6:53
136.5 C 224kPa 7:23
136.5 C 224kPa 7:23
VENTING 7:23
AIR DRYING 8:14
CYCLE COMPLETE 8:18

Digital Signature #
604270277957DE00
