
Midmark M11 - v1.0.5
Total Cycles: 568

Name: RITTER2
Sterilizer ID: V2313228
11/18/2021 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 170.2 F 3.3
4:00 190.0 F 8.3
6:00 214.1 F 11.0
8:00 229.0 F 9.4
10:00 235.1 F 9.3
12:00 243.6 F 13.2
14:00 243.9 F 13.1
16:00 257.2 F 20.1
18:00 268.7 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.5 F 29.5
2:00 272.7 F 30.0
3:00 271.6 F 29.4
4:00 272.1 F 29.6

Min 271.3 F 29.2
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 224.8 F 0.0
10:00 207.6 F 0.0
15:00 211.9 F 0.0
20:00 208.0 F 0.0
25:00 219.7 F 0.0
Dry Cycle Aborted

FILLING: 2:56
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 26:25
TOTAL CYCLE: 00:55:36

11/18/2021 04:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____