
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002888
11:03 21/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:43
STERILIZING 3:48
136.5 C 224kPa 3:48
136.4 C 225kPa 4:19
136.4 C 224kPa 4:49
136.4 C 224kPa 5:19
136.3 C 224kPa 5:49
136.3 C 223kPa 6:19
136.4 C 224kPa 6:49
136.4 C 223kPa 7:19
136.4 C 223kPa 7:19
VENTING 7:19
AIR DRYING 8:12
CYCLE COMPLETE 8:14

Digital Signature #
7C5434B230DD9AF7
