

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2007

Name: RITTER2  
Sterilizer ID: V2313228  
02/03/2023 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	80.8 F	0.0
2:00	82.5 F	0.0
4:00	93.5 F	0.0
6:00	127.8 F	0.0
8:00	177.5 F	0.2
10:00	202.3 F	0.7
12:00	217.9 F	4.3
14:00	228.5 F	8.1
16:00	239.2 F	12.6
18:00	238.2 F	10.1
20:00	236.4 F	8.4
22:00	239.9 F	9.8
24:00	250.0 F	15.2
26:00	263.0 F	22.7

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 254F  
CHAMBER PRESS 17.3PSI  
\*\* ITEMS NOT STERILE \*\*

02/03/2023 02:43 PM