
Midmark M11 - v1.0.5
Total Cycles: 3685

Name: RITTER2
Sterilizer ID: V2313228
06/11/2024 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.1 F 0.0
2:00 98.0 F 0.3
4:00 118.5 F 2.0
6:00 150.8 F 4.7
8:00 169.5 F 7.6
10:00 188.7 F 11.1
12:00 204.5 F 4.6
14:00 215.3 F 5.9
16:00 224.8 F 9.6
18:00 235.0 F 11.7
20:00 236.7 F 10.5
22:00 237.6 F 8.9
24:00 247.0 F 14.2
26:00 257.7 F 20.0
28:00 267.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 273.2 F 29.7
2:00 271.9 F 29.0
3:00 272.2 F 29.6
4:00 272.9 F 29.6

Min 271.5 F 28.9
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 199.6 F 0.0
10:00 202.1 F 0.0
15:00 199.1 F 0.0
20:00 200.4 F 0.0
25:00 199.4 F 0.0
30:00 196.6 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:30
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

06/11/2024 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____