
Midmark M11 - v1.0.5
Total Cycles: 3198

Name: RITTER2
Sterilizer ID: V2313228
12/08/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.4 F 0.0
2:00 146.7 F 1.4
4:00 157.3 F 4.4
6:00 181.9 F 8.3
8:00 200.9 F 12.9
10:00 218.4 F 7.0
12:00 228.5 F 11.9
14:00 231.7 F 8.9
16:00 239.1 F 10.6
18:00 245.0 F 13.6
20:00 257.0 F 20.0
22:00 267.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.8 F 30.2
2:00 272.5 F 29.8
3:00 271.4 F 29.3
4:00 271.4 F 29.7

Min 271.1 F 29.2
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 224.5 F 0.0
10:00 214.2 F 0.0
15:00 205.3 F 0.0
20:00 201.0 F 0.0
25:00 199.5 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:02
CYCLE COMPLETE

12/08/2023 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____