
Midmark M11 - v1.0.5
Total Cycles: 3627

Name: RITTER2
Sterilizer ID: V2313228
05/16/2024 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.4 F 0.0
2:00 118.2 F 0.9
4:00 128.4 F 3.4
6:00 154.9 F 6.6
8:00 179.6 F 10.4
10:00 207.7 F 6.5
12:00 215.0 F 7.3
14:00 226.2 F 11.6
16:00 227.8 F 7.4
18:00 239.4 F 12.6
20:00 238.9 F 10.8
22:00 250.8 F 16.6
24:00 261.3 F 22.7
26:00 271.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.4 F 30.0
2:00 271.2 F 29.3
3:00 272.4 F 30.2
4:00 271.8 F 29.6

Min 271.0 F 29.3
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 203.2 F 0.0
10:00 201.5 F 0.0
15:00 198.2 F 0.0
20:00 197.2 F 0.0
25:00 210.5 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:42
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:05:17
CYCLE COMPLETE

05/16/2024 02:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____