
Midmark M11 - v1.0.0
Total Cycles: 3147

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 127.3 F 0.0
4:00 152.7 F 0.0
6:00 191.8 F 0.0
8:00 206.0 F 4.4
10:00 219.5 F 8.5
12:00 234.4 F 12.9
14:00 234.4 F 9.1
16:00 239.7 F 10.4
18:00 244.5 F 12.7
20:00 243.2 F 11.7
22:00 256.5 F 18.6
24:00 268.1 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.2 F 28.5
2:00 272.3 F 28.9
3:00 271.8 F 28.2
4:00 272.6 F 28.9

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 213.1 F 0.0
10:00 229.1 F 0.0
15:00 240.9 F 0.0
20:00 231.3 F 0.0
25:00 224.3 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:03:26
CYCLE COMPLETE

04/21/2023 04:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____