

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4356

Name: RITTER2  
Sterilizer ID: V2313228  
02/19/2025 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 82.1 F 0.2  
4:00 125.2 F 0.9  
6:00 102.4 F 2.4  
8:00 143.8 F 5.7  
10:00 193.0 F 9.4  
12:00 186.6 F 13.1  
14:00 205.7 F 3.8  
16:00 216.0 F 7.3  
18:00 226.7 F 11.0  
20:00 228.5 F 7.2  
22:00 239.7 F 11.5  
24:00 237.0 F 9.3  
26:00 249.9 F 15.5  
28:00 260.8 F 21.7  
30:00 270.9 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.2 F 29.0  
2:00 272.7 F 29.5  
3:00 271.8 F 29.0  
4:00 271.8 F 28.8

Min 271.5 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.6 F 0.0  
5:00 189.4 F 0.0  
10:00 181.7 F 0.0  
15:00 186.3 F 0.0  
20:00 190.9 F 0.0  
25:00 193.2 F 0.0  
30:00 202.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:41  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:09:04  
CYCLE COMPLETE

02/19/2025 10:53 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_