
Midmark M11 - v1.0.0
Total Cycles: 1819

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2022 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 173.7 F 3.1
4:00 193.5 F 7.0
6:00 224.2 F 12.4
8:00 223.4 F 6.3
10:00 236.3 F 11.2
12:00 232.8 F 7.7
14:00 245.1 F 12.4
16:00 244.4 F 12.3
18:00 257.0 F 19.0
20:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.7
2:00 271.1 F 28.2
3:00 272.4 F 28.9
4:00 271.4 F 28.2

Min 271.1 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 228.1 F 0.0
10:00 204.1 F 0.0
15:00 205.7 F 0.0
20:00 227.5 F 0.0
25:00 236.4 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

04/13/2022 12:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____