
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000389
15:17 01/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:34
STERILIZING 4:42
136.0 C 220kPa 4:42
136.0 C 221kPa 5:12
136.3 C 224kPa 5:42
136.4 C 223kPa 6:12
136.1 C 219kPa 6:42
136.4 C 223kPa 7:12
136.3 C 224kPa 7:42
136.4 C 224kPa 8:12
136.4 C 224kPa 8:12
VENTING 8:12
AIR DRYING 9:07
CYCLE COMPLETE 9:09

Digital Signature #
69007D11B89D57F1
