
Midmark M11 - v1.0.5
Total Cycles: 1899

Name: RITTER2
Sterilizer ID: V2313228
01/05/2023 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 186.6 F 3.2
4:00 193.3 F 7.9
6:00 212.2 F 13.3
8:00 225.6 F 8.6
10:00 235.9 F 11.1
12:00 240.9 F 12.2
14:00 240.4 F 11.1
16:00 253.5 F 17.8
18:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.7
2:00 271.6 F 29.0
3:00 272.3 F 29.9
4:00 272.4 F 29.5

Min 271.4 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 243.8 F 0.0
10:00 226.7 F 0.0
15:00 224.3 F 0.0
20:00 224.9 F 0.0
25:00 230.0 F 0.0
30:00 218.5 F 0.0

Drying Complete

FILLING: 1:57
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

01/05/2023 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____