
Midmark M11 - v1.0.0
Total Cycles: 1941

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2022 12:45 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.1 F 0.0
2:00 191.1 F 1.7
4:00 200.0 F 6.6
6:00 221.2 F 12.8
8:00 230.5 F 8.7
10:00 236.1 F 8.9
12:00 245.8 F 12.8
14:00 245.1 F 12.8
16:00 259.2 F 20.3
18:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.0 F 28.3
2:00 272.7 F 28.8
3:00 272.4 F 28.9

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
Dry Cycle Aborted

FILLING: 2:24
HEATING: 18:38
STERILIZING: 3:00
VENTING: 2:46
DRYING: 0:08
TOTAL CYCLE: 00:26:56

05/16/2022 01:11 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____