

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001517  
15:42 10/20/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:40  
PRESSURIZING 5:22  
STERILIZING 6:52  
136.0 C 220kPa 6:52  
136.0 C 219kPa 7:22  
136.4 C 223kPa 7:52  
136.5 C 223kPa 8:22  
136.4 C 223kPa 8:52  
136.3 C 222kPa 9:22  
136.4 C 223kPa 9:52  
136.3 C 223kPa 10:22  
VENTING 10:22  
AIR DRYING 11:27  
CYCLE COMPLETE 41:27

-----  
Digital Signature #  
BE327AA93F5AD2A3  
-----