

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4256

Name: RITTER 1  
Sterilizer ID: V2178198  
02/01/2024 09:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 86.7 F 0.2  
4:00 108.4 F 1.0  
6:00 111.4 F 3.0  
8:00 136.9 F 5.7  
10:00 168.5 F 9.0  
12:00 187.8 F 12.6  
14:00 208.8 F 3.4  
16:00 216.4 F 6.7  
18:00 226.1 F 10.2  
20:00 232.7 F 8.8  
22:00 237.4 F 9.9  
24:00 237.8 F 9.3  
26:00 245.2 F 12.9  
28:00 256.6 F 18.9  
30:00 266.7 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 272.6 F 29.1  
2:00 271.7 F 28.2  
3:00 272.3 F 29.0  
4:00 271.9 F 28.6

Min 271.0 F 28.1  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
5:00 238.2 F 0.0  
10:00 212.2 F 0.0  
15:00 221.2 F 0.0  
20:00 218.8 F 0.0  
25:00 221.8 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:42  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:09:53  
CYCLE COMPLETE

02/01/2024 10:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_