
Midmark M11 - v1.0.0
Total Cycles: 5192

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2024 11:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 160.8 F 0.0
4:00 172.6 F 1.0
6:00 189.7 F 4.5
8:00 206.3 F 8.2
10:00 219.9 F 12.2
12:00 225.6 F 6.7
14:00 236.6 F 11.8
16:00 235.7 F 9.0
18:00 240.9 F 10.3
20:00 247.1 F 13.6
22:00 258.7 F 19.8
24:00 268.7 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.0 F 28.6
2:00 272.4 F 28.6
3:00 271.9 F 28.2
4:00 271.3 F 27.9

Min 271.2 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 211.7 F 0.0
10:00 201.9 F 0.0
15:00 210.0 F 0.0
20:00 226.4 F 0.0
25:00 236.8 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:21
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:04:03
CYCLE COMPLETE

12/16/2024 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____