
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000875
15:04 18/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:19
PRESSURIZING 4:20
STERILIZING 5:49
136.3 C 231kPa 5:49
136.2 C 230kPa 6:19
136.2 C 231kPa 6:49
136.4 C 233kPa 7:19
136.2 C 229kPa 7:49
136.1 C 229kPa 8:19
136.1 C 229kPa 8:49
136.3 C 231kPa 9:19
136.3 C 231kPa 9:19
VENTING 9:19
AIR DRYING 10:17
CYCLE COMPLETE 25:18

Digital Signature #
732379AC91AED02A
