
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007432
9:12 27/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:20
136.0 C 220kPa 3:20
136.3 C 220kPa 3:50
136.1 C 218kPa 4:20
136.4 C 225kPa 4:50
136.4 C 223kPa 5:20
136.1 C 222kPa 5:50
136.3 C 220kPa 6:20
136.4 C 224kPa 6:50
136.4 C 224kPa 6:50
VENTING 6:50
AIR DRYING 7:39
CYCLE COMPLETE 7:44

Digital Signature #
D5843BE9B7FE77F6
