
Midmark M11 - v1.0.5
Total Cycles: 2330

Name: RITTER2
Sterilizer ID: V2313228
05/03/2023 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.7 F 0.0
2:00 195.3 F 3.5
4:00 204.1 F 9.3
6:00 223.2 F 7.7
8:00 234.6 F 11.7
10:00 236.4 F 10.1
12:00 236.0 F 9.2
14:00 251.6 F 16.7
16:00 264.3 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 271.8 F 29.1
2:00 272.5 F 29.5
3:00 271.2 F 29.1
4:00 271.9 F 29.2

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 215.8 F 0.0
10:00 223.5 F 0.0
15:00 223.4 F 0.0
20:00 223.2 F 0.0
25:00 224.3 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:21
CYCLE COMPLETE

05/03/2023 11:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____