

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2198

Name: RITTER2  
Sterilizer ID: V2313228  
03/30/2023 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 217.9 F 0.0  
2:00 211.1 F 4.3  
4:00 210.8 F 10.3  
6:00 224.1 F 6.5  
8:00 236.5 F 13.0  
10:00 240.1 F 11.4  
12:00 240.2 F 10.7  
14:00 254.7 F 18.1  
16:00 266.8 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.2 F 29.3  
2:00 271.3 F 28.9  
3:00 272.2 F 29.5  
4:00 271.6 F 28.9  
  
Min 271.2 F 28.7  
Max 272.3 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
5:00 226.5 F 0.0  
10:00 236.1 F 0.0  
15:00 229.3 F 0.0  
20:00 227.5 F 0.0  
25:00 226.8 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 2:34  
HEATING: 17:27  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:16  
CYCLE COMPLETE

03/30/2023 11:53 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_