
Midmark M11 - v1.0.5
Total Cycles: 3512

Name: RITTER2
Sterilizer ID: V2313228
04/10/2024 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.5 F 0.0
2:00 122.0 F 0.1
4:00 118.6 F 0.9
6:00 145.5 F 3.8
8:00 172.5 F 7.2
10:00 193.6 F 11.0
12:00 217.3 F 5.4
14:00 222.6 F 9.1
16:00 232.6 F 12.8
18:00 234.3 F 10.5
20:00 235.8 F 8.6
22:00 246.8 F 14.4
24:00 257.8 F 20.4
26:00 268.0 F 26.9
28:00 271.9 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 272.1 F 29.7
2:00 272.6 F 29.9
3:00 272.7 F 30.1
4:00 272.3 F 30.3

Min 271.6 F 29.7
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 217.7 F 0.0
10:00 217.5 F 0.0
15:00 214.2 F 0.0
20:00 212.9 F 0.0
25:00 204.4 F 0.0
30:00 200.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:04
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:06:36
CYCLE COMPLETE

04/10/2024 11:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____