
Midmark M11 - v1.0.0
Total Cycles: 3681

Name: RITTER 1
Sterilizer ID: V2178198
08/21/2023 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 174.5 F 3.0
4:00 175.7 F 6.8
6:00 193.6 F 11.2
8:00 217.2 F 4.8
10:00 226.2 F 9.8
12:00 234.9 F 8.6
14:00 241.3 F 11.8
16:00 239.0 F 10.0
18:00 251.8 F 16.4
20:00 263.3 F 22.9
22:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.2 F 28.7
2:00 272.7 F 29.2
3:00 271.7 F 28.7
4:00 272.3 F 28.8

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 224.8 F 0.0
10:00 237.6 F 0.0
15:00 231.9 F 0.0
20:00 223.6 F 0.0
25:00 225.0 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

08/21/2023 02:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____