
Midmark M11 - v1.0.5
Total Cycles: 1991

Name: RITTER2
Sterilizer ID: V2313228
02/01/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 154.8 F 0.0
4:00 166.7 F 2.9
6:00 191.1 F 7.5
8:00 219.6 F 12.7
10:00 223.9 F 7.8
12:00 236.0 F 13.2
14:00 237.9 F 10.7
16:00 236.5 F 9.2
18:00 250.6 F 16.2
20:00 262.8 F 23.2
22:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 29.8
2:00 271.6 F 29.0
3:00 272.5 F 29.9
4:00 272.2 F 29.4

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 207.0 F 0.0
10:00 219.7 F 0.0
15:00 217.9 F 0.0
20:00 216.7 F 0.0
25:00 223.5 F 0.0
30:00 232.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

02/01/2023 10:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____