
Midmark M11 - v1.0.5
Total Cycles: 388

Name: RITTER2
Sterilizer ID: V2313228
09/23/2021 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 170.0 F 2.9
4:00 192.3 F 8.0
6:00 211.4 F 12.1
8:00 225.5 F 8.9
10:00 235.0 F 10.8
12:00 241.0 F 12.5
14:00 241.4 F 11.7
16:00 254.5 F 18.5
18:00 266.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.3 F 29.4
2:00 272.6 F 29.8
3:00 272.1 F 29.8
4:00 272.2 F 29.3

Min 271.7 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 221.7 F 0.0
10:00 221.8 F 0.0
15:00 240.4 F 0.0
20:00 218.6 F 0.0
25:00 218.9 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

09/23/2021 09:59 AM
ACCEPTED: X
REJECTED:
Operator: U02