
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001297
16:09 13/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 3:00
STERILIZING 4:11
136.4 C 222kPa 4:11
136.2 C 224kPa 4:41
136.3 C 224kPa 5:11
136.3 C 224kPa 5:41
136.2 C 225kPa 6:11
136.4 C 226kPa 6:41
136.2 C 223kPa 7:11
136.4 C 225kPa 7:41
136.4 C 225kPa 7:41
VENTING 7:41
AIR DRYING 8:33
CYCLE COMPLETE 12:00

Digital Signature #
70FF9F4BFF45C69E
