

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001396  
18:23 22/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:02  
PRESSURIZING 4:04  
STERILIZING 5:20  
136.4 C 224kPa 5:20  
136.4 C 224kPa 5:50  
136.3 C 223kPa 6:20  
136.2 C 222kPa 6:50  
136.3 C 225kPa 7:20  
136.3 C 224kPa 7:50  
136.3 C 224kPa 8:20  
136.3 C 224kPa 8:50  
136.3 C 224kPa 8:50  
VENTING 8:50  
AIR DRYING 9:45  
CYCLE COMPLETE 24:45  
-----

Digital Signature #  
74019FB2EE1E596B  
-----