
Midmark M11 - v1.0.0
Total Cycles: 2545

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 06:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.8 F 0.0
2:00 159.9 F 2.8
4:00 175.6 F 6.3
6:00 199.4 F 11.2
8:00 215.3 F 4.6
10:00 226.5 F 9.1
12:00 237.4 F 12.6
14:00 238.2 F 10.4
16:00 235.4 F 8.1
18:00 249.8 F 15.1
20:00 261.4 F 21.7
22:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.6 F 29.1
2:00 271.7 F 28.3
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 218.7 F 0.0
10:00 221.3 F 0.0
15:00 224.8 F 0.0
20:00 225.8 F 0.0
25:00 226.1 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:22
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:52
CYCLE COMPLETE

11/10/2022 07:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____