
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004925
11:11 09/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:12
STERILIZING 4:12
136.4 C 222kPa 4:12
136.2 C 222kPa 4:42
CYCLE ABORTED 4:58
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:58

Digital Signature #
228735AB96EBFCA9
