
Midmark M11 - v1.0.5
Total Cycles: 3592

Name: RITTER2
Sterilizer ID: V2313228
05/06/2024 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.7 F 0.0
2:00 130.7 F 1.3
4:00 140.7 F 3.6
6:00 165.2 F 6.7
8:00 185.2 F 10.6
10:00 210.4 F 6.1
12:00 217.2 F 7.6
14:00 230.6 F 11.8
16:00 229.0 F 7.8
18:00 241.0 F 12.9
20:00 239.8 F 11.0
22:00 251.3 F 16.5
24:00 261.1 F 22.3
26:00 270.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.4 F 30.1
2:00 272.4 F 29.7
3:00 272.1 F 30.0
4:00 272.8 F 29.9

Min 271.0 F 28.9
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 213.4 F 0.0
10:00 208.5 F 0.0
15:00 200.4 F 0.0
20:00 212.4 F 0.0
25:00 217.6 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:52
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:30
CYCLE COMPLETE

05/06/2024 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____