

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001988  
8:30 09/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:06  
STERILIZING 3:07  
136.4 C 223kPa 3:07  
136.0 C 223kPa 3:37  
136.4 C 223kPa 4:07  
136.4 C 225kPa 4:37  
136.4 C 224kPa 5:07  
136.2 C 223kPa 5:37  
136.3 C 224kPa 6:07  
136.0 C 225kPa 6:37  
136.0 C 225kPa 6:37  
VENTING 6:37  
AIR DRYING 7:27  
CYCLE COMPLETE 7:45

-----  
Digital Signature #  
3DB5096268F6FDE4  
-----