
Midmark M11 - v1.0.5
Total Cycles: 2219

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 07:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.9 F 0.0
2:00 81.7 F 0.0
4:00 103.8 F 0.0
6:00 123.9 F 0.1
8:00 161.1 F 1.5
10:00 173.4 F 5.1
12:00 191.4 F 8.9
14:00 208.0 F 12.7
16:00 220.3 F 6.2
18:00 229.2 F 10.4
20:00 234.2 F 8.4
22:00 241.6 F 12.2
24:00 240.1 F 10.6
26:00 252.3 F 16.7
28:00 262.9 F 23.0
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.7 F 29.2
2:00 271.4 F 29.0
3:00 272.3 F 29.0
4:00 271.6 F 29.2

Min 271.2 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 221.3 F 0.0
10:00 212.1 F 0.0
15:00 211.8 F 0.0
20:00 214.3 F 0.0
25:00 213.2 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:18
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

04/10/2023 08:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____