

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 004225  
9:23 24/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:27  
STERILIZING 3:18  
136.0 C 220kPa 3:18  
136.3 C 223kPa 3:48  
136.3 C 223kPa 4:18  
136.3 C 221kPa 4:48  
136.4 C 224kPa 5:18  
136.4 C 223kPa 5:48  
136.3 C 222kPa 6:18  
136.4 C 225kPa 6:48  
136.4 C 225kPa 6:48  
VENTING 6:48  
AIR DRYING 7:35  
CYCLE COMPLETE 7:42

-----  
Digital Signature #  
60F4C99193F10EF7  
-----