
Midmark M11 - v1.0.5
Total Cycles: 953

Name: RITTER2
Sterilizer ID: V2313228
04/04/2022 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 173.7 F 1.1
4:00 202.3 F 6.9
6:00 221.7 F 13.3
8:00 232.0 F 9.8
10:00 234.7 F 8.6
12:00 240.7 F 10.7
14:00 248.5 F 14.9
16:00 261.9 F 22.4
18:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.9 F 29.5
2:00 271.7 F 28.7
3:00 273.0 F 29.7
4:00 272.3 F 29.0

Min 271.6 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 209.7 F 0.0
10:00 221.5 F 0.0
15:00 229.2 F 0.0
20:00 229.1 F 0.0
25:00 229.5 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 18:15
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:56:39
CYCLE COMPLETE

04/04/2022 02:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____